258, Montée des Grands Champs
71520 NAVOUR sur Grosne
FRANCE

# MODEL-LOCK CUP PLATE

## Reference 590050

## INSTRUCTIONS FOR USE

**1 — Model preparation**
Trim the base of each of the two printed resin models so that it is flat.

**IMPORTANT:** The models must have a **perfectly flat base** to ensure optimal bonding.

**2 — Cleaning**
Dry the models and carefully remove any dust before bonding. The surfaces must be clean and dry.

**3 — Bonding the models**
Using a **cyanoacrylate adhesive**, bond the models to the **smooth side** of the **Model-Lock PLATE Ref. 590050**.

**4 — Snap-fitting**
On the side opposite the model, snap the **black female Model-Lock CUP component Ref. 493200** into place.

Check that it is properly engaged and that the assembly is stable.

**5 — Bonding the occlusal posts**
Bond the **occlusal posts** using a cyanoacrylate adhesive.

**6 — Mounting on the articulator**
Mount on the articulator using the conventional method.

**7 — Releasing the occlusion**
Once the plaster has fully set, cut the occlusal posts to release the occlusion.

**8 — Removal**
Unclip the model together with its **Model-Lock CUP PLATE** from the articulator mounting.

The Model-Lock system allows **precise and repeatable repositioning of the model on the articulator**.

**PRECAUTIONS**

Product intended for **professional use in dental laboratories**.

Use a suitable cyanoacrylate adhesive and follow the adhesive manufacturer’s recommendations.

Do not unclip until all materials have fully set.